

Government of West Bengal
Directorate of Health Services
Central Medical Stores
141, A. J.C. Bose Road, Kolkata-700014.

Memo. No. HST/1P-09-2014/Pt-1/742

Dated, Kolkata, the 21/08/2014.

NOTICE

All concerned are hereby informed that the examination and selection of the item under the e-TENDER FOR PROCUREMENT OF MEDICAL DEVICES, SURGICAL INSTRUMENTS, HOSPITAL EQUIPMENTS & OTHER EQUIPMENTS UNDER JSSK AND OTHERS FOR THE YEAR 2014-2016 Vide NIT No. HST/1P-09-2014/Pt-1/22, dated : 01.04.2014 will be held as per schedule given below at the office chamber of the undersigned.

Approved Tenderers are requested to produce their samples as per the direction of the Terms & Conditions and bring all the related papers (including drug license & Drug Endorsement list where applicable and brochures). They are also requested to mention the name of the firm, Cat. No. and name of the item on the TAG attached to the sample only.

Sl. No.	Name of the items with specification	Cat. No	Name of the company	Date & Time of Samples
1.	<u>Single blood bag 350 ml</u> Specification:- DEHP Plastic, sterile, pyrogen free, filled with clear 49 ml CPDA solution, closed systems, transparent container complete with flexible non-kinking, non-sticking collection tube, impregnated with unique multiple ID/Segment number. The needle should be of 16 gauge, ultra-thin sharp, bevelled tip. 2 pairs of aluminium metal clip should accompany each bag. The specification should be as per ISO standard. The detail of the specification is made available in the tender document and NACO website.	BT 32	1.HILL LIFE CARE LTD. 2.MITRA INDUSTRIES PVT. LTD. 3.POLY MEDICURE LTD.	27.08.2014 AT 12.00 NOON
2.	<u>Double blood bag 350 ml</u> Specification:- One primary bag (350 ml) and one satellite bag (300 ml). DEHP/TOTM Plastic, sterile, pyrogen free, filled with clear 49 ml CPDA solution for primary bag, closed system, transparent container complete with flexible non-kinking, non-sticking collection tube, impregnated with unique multiple ID/Segment number. The needle should be of 16 gauge, ultra-thin sharp, bevelled tip. 2 pairs of aluminium metal clip should accompany each bag. The specification should be as per ISO standard. The detail of the specification is made available in the tender document and NACO website.	BT 33	1.HILL LIFE CARE LTD. 2.MITRA INDUSTRIES PVT. LTD. 3.POLY MEDICURE LTD.	27.08.2014 AT 1200 NOON
3.	<u>Triple blood bag 350 ml</u> Specification:- One primary bag (350 ml) and one 1st satellite bag (300 ml) & one 2nd satellite bag (300 ml) for platelet storage for 5 days. DEHP/TOTM Plastic, sterile, pyrogen free, filled with clear 49 ml CPDA solution for primary bag, closed system, transparent container complete with flexible non-kinking, non-sticking collection tube, impregnated with unique multiple ID/Segment number. The needle should be of 16 gauge, ultra-thin sharp, bevelled tip. 3 pairs of aluminium metal clip should accompany each bag. The specification should be as per ISO standard. The detail of the specification is made available in the tender document and NACO website.	BT 34	1.HILL LIFE CARE LTD. 2.MITRA INDUSTRIES PVT. LTD. 3.POLY MEDICURE LTD.	27.08.2014 AT 12.00 NOON

Contd at page - 2.

