



Government of West Bengal
Office of the Principal
College of Medicine & Sagore Dutta Hospital
578, B.T.Road, Kamarhati, Kolkata-700058

Memo No: CMSDH/Pr./2045/21

Date: 26-10-2021

Corrigendum Tender Notice

NIT No: CMSDH-13/2021-22

Tender ID: 2021_HFW_346990_1

Tender Name: Supply, Installation and Commissioning of Video Bronchoscope

Ref: Pre bid meeting dated 26.10.2021

Following points in technical specification are hereby modified:-

SI No	SI No in Specification	Existing	Modification
1	1	The scope should have High definition image quality with one touch water-proof connector	The scope should have High definition image quality desirable with one touch water-proof connector
2	1	It should have real time optical image enhancement technology that improves the visualization of vessels on the mucosal surface (Real time Narrow Band Imaging)	It should have real time optical image enhancement technology that improves the visualization of vessels on the mucosal surface (Real time Narrow Band Imaging- NBI/BLI/OE i-Scan)
3	1	Depth of field in mm 3-100 mm	Depth of field in mm 2-100 mm or more
4	1	Compatible with leakage testing device with its air flow and pressure regulation through light source's air pump	Compatible with leakage testing device with its air flow and pressure regulation through light source's air pump (Desirable)/Manual Leakage Testing
5	3	Should be equipped with special detection technology for detailed observation by enhancing visibility of blood capillaries or mucosa (REAL TIME NARROW BAND IMAGING)	Should be equipped with special detection technology for detailed observation by enhancing visibility of blood capillaries or mucosa (REAL TIME NARROW BAND IMAGING- NBI/BLI/OE i-Scan)

Dhs 26/10/2021

Principal
College of Medicine & Sagore Dutta Hospital
Kamarhati, Kolkata-700058