



सत्यमेव जयते

Government of West Bengal

Department of Health and Family Welfare

Office of the Principal, Burdwan Medical College, Baburbag, Burdwan - 713104

Phone no- (0342) 7962201

Mail ID- burdwanmedicalcollege76@gmail.com

Memo No: BMC/ 2208

Dated, Burdwan, the 7th July, 2022

CORRIGENDUM

Ref.: This office vide NIT Memo No. BMC/1715 dated 01.06.2022
e-Tender ID: 2022_HFW_383551_1

For the above mentioned reference and e-Tender ID, the revised technical specification as per resolution taken in the pre-bid meeting held on 16.06.2022 for the equipments "OCT Machine, Autorefractometer, Non-contact Tonometer (NCT) Apparatus and Slit Lamp" for the Department of Ophthalmology, is attached herewith.

The other things for the said e-tender will remain unchanged.

Principal
Burdwan Medical College
Burdwan

To
The Principal
Burdwan Medical College
Purba Bardhaman
Pin-713104



Ref Tender No: BMC/1715 **Date:** 01/06/2022

Sub : Requesting for Amendment of the Technical Specification for Autorefractometer

Respected Sir,

With due respect, we want to draw your kind attention towards an important point that some necessary Amendment need to be done in the existing Technical Specification of some points so as to allow wider participation and for more effective and better service for the patients.

So, We would request you to make the following amendments in tender terms and Specifications :-

SERIAL DETAILS (AS PER TENDER)	PRESENT SPECS	SPECS TO BE AMENDED	REMARK
As Per Page No. 30, For Amendment Point " Motorised chinrest movement range 65 mm" in Technical Specification	Motorised chinrest movement range 65 mm	Motorised chinrest movement range 65 mm or manual	For wider participation
As Per Page No. 30, For Amendment Point " Tilttable 6.4 Inch TFT display" in Technical Specification	Tilttable 6.4 Inch TFT display	Tilttable 6.4 Inch TFT display or Standard	For wider participation
As Per Page No. 30, For Amendment Point " CE/FDA Approved" in Technical Specification	CE/ FDA Approved	CE and FDA Approved	For international Standard quality product and better patient care

Thanking you in advance

Date:-

Mousumi Bandyopadhyay
7/7/22
HOD & Professor
Dept. of Ophthalmology, BMCH
Professor & Head
Department of Ophthalmology
Burdwan Medical College & Hospital
Burdwan - 713104

To
The Principal
Burdwan Medical College
Purba Bardhaman
Pin-713104



Ref Tender No: BMC/1715 **Date:** 01/06/2022

Sub : Requesting for Amendment of the Technical Specification for Non Contact Tonometer (NCT) Apparatus.

Respected Sir,

With due respect, we want to draw your kind attention towards an important point that some necessary Amendment need to be done in the existing Technical Specification of some points so as to allow wider participation and for more effective and better service for the patients.

So, We would request you to make the following amendments in tender terms and Specifications :-

SERIAL DETAILS (AS PER TENDER)	PRESENT SPECS	SPECS TO BE AMENDED	REMARK
As Per Page No. 31, For Amendment Point No.2 in Technical Specification	Auto- Tracking to find the right position and initiate the measurement	Auto- Tracking to find the right position and initiate the measurement or Auto Shoot	For wider participation
As Per Page No. 31, For Amendment Point No.6 in Technical Specification	Monitor 5.7" LCD TFT	Monitor 5.7" LCD TFT or more	For better patient care
As Per Page No. 31, For Amendment Point No. 9 in Technical Specification	Power consumption 60-85 VA	Power consumption 60-85 VA or As per Standrd	For wider participation
As Per Page No. 31, For Amendment Point No. 16 in Technical Specification	CE/ FDA Approved	CE and FDA Approved	For international Standard quality product and better patient care

Thanking you in advance

Mousumi Banerjee
7/17/22
PROF. & HOD
Dept. of Ophthalmology, BMCH
Professor & Head
Department of Ophthalmology
Burdwan Medical College & Hospital
Burdwan - 713104

To
The Principal
Burdwan Medical College
Purba Bardhaman
Pin-713104



Ref Tender No: BMC/1715 **Date:** 01/06/2022

Sub : Requesting for Amendment of the Technical Specification for SLIT Lamp

Respected Sir,

With due respect, we want to draw your kind attention towards an important point that some necessary Amendment need to be done in the existing Technical Specification of some points so as to allow wider participation and for more effective and better service for the patients.

So. We would request you to make the following amendments in tender terms and Specifications :-

SERIAL DETAILS (AS PER TENDER)	PRESENT SPECS	SPECS TO BE AMENDED	REMARK
As Per Page No. 32, For Amendment Point " Light Source" in Technical Specification	Light Source : Halogen Light	Light Source: Halogen Light/ LED	For wider participation
As Per Page No. 32, For Amendment Point " Eyepiece magnification" in Technical Specification	Eyepiece Magnification : Super High eyepoint eyepieces 10X with Compensation of ametropia +- 8D	Eyepiece Magnification : Super High eyepoint eyepieces 10X/12.5 X with Compensation of ametropia +- 8D	For wider participation
As Per Page No. 32, For Amendment Point" Light Source" in Technical Specification	Light Source: 6V 10W halogen lamp	Light Source: 6V 10W halogen/Led lamp	For wider participation

Thanking you in advance

DATE:-

Moreswar Bandyopadhyay
PROF. & HOD 7/7/22

Dept. of Ophthalmology, BMCH
Department of Ophthalmology
Burdwan Medical College & Hospital
Burdwan - 713104