

**Government of West Bengal**  
**Office of the Chief Medical Officer of Health**  
**5 D. L. Roy Road, Krishnanagar, Nadia**

Telephone: (03472)252306

Email ID: cmoh\_nad@wbhealth.gov.in/ cmohnadia@gmail.com

Memo No. CMOH-Nad/ 2528

Dated. Krishnanagar the March 14, 2019

To  
The Proprietor,  
Aisha Enterprise  
Sajanopally, Suri, Birbhum

This is to inform you that as per decision taken by the tender selection committee in connection with re-tender for procurement of Bio-safety cabinet vide this office Memo No.-CMOH-Nad/9102, dt.-05/10/2018, you have been accepted with their quoted rate i.e. Rs.1, 89, 786/- (one lakh eighty nine thousand seven hundred and eighty six) for per Biosafety cabinet. This rate for per Biosafety cabinet would remain valid for 1(one) year and may be extended, if required as per terms & conditions of tender. In case of any discrepancy within the validity period, the authority reserves the right to take disciplinary action as per norms.

You are requested to supply & install one (1) pc bio-safety cabinet @ Rs.1 , 89, 786/- (one lakh eighty nine thousand seven hundred and eighty six) (including all applicable taxes and charges and A.M.C cost for 1 (one ) year, replacement cost of parts for damage done on the normal course of functioning & ware and tear.) within 5 days at the District Public Health Laboratory, Shaktinagar Campus, District Hospital, Nadia. The specification of the bio-safety cabinet will be as below:-

TECHNICAL SPECIFICATION

Class-II A2 Biosafety Cabinet (4 ft)

1. The cabinet should be advanced microprocessor control, which supervises & intervenes operation of all cabinet functions. Temperature-compensated air velocity sensor monitors both exhaust and down flow. 24-hour clock, UV timer, UV run hour meter, and blower run hour meter.
2. There should be provision of an indicator for filter change (when required).
3. There should be safety interlocking system for protection of instrument & user.
4. The cabinet should have high efficient blower motor preferably DC.
5. The cabinet should have long life DUAL ULPA/ HEPA Filter for supply and exhaust (per IEST-RP-CC001.3) with 99.999% efficiency for particle size at least up to 0.3 microns.
6. Should have raised armrest for elevating the operator's arms to prevent inflow grille blockage for safety work.
7. Work tray should be made of single piece stainless steel type 304.
8. Programmable automatic UV light timer should simplify operation and extending UV light life and saving energy.
9. There should be sliding front sash (slightly slanted) which can be fully opened to insert and remove large instruments.

10. The cabinet should come with following accessories: UV lamp, electrical outlet sockets and fitted with castor wheels for easy movability. Front wheels lockable with brakes [antimicrobial coated] .

11. Should be provided original literature.


12. Certifications:

Product certification: CE; Should be supported with documentations or US FDA (The product must be enlisted under US FDA and the quoted model must have US FDA permission to market in US & export to other countries.

Quality certification: ISO certified

After successful delivery & installation of the bio-safety cabinet, you may submit the bill to this office along with the Consignee Receipt Certificate and Satisfactory Installation Certificate

You are requested to submit an additional amount of Rs.15000/-(fifteen thousand) as the performance security to this office within 3 working days of this Memo


  
**Chief Medical Officer of Health**  
Nadia S. Chanda

**Memo No. CMOH-Nad/ 2528/1(8) .**

**Dated. Krishnagar the March 14, 2019**

Copy forwarded for information & taking necessary action for wide publication through their notice-board to:

- 1) The Sabhadhipati, Nadia Zilla Parishad
- 2) The District Magistrate, Nadia
- 3) The District Informatics Officer, N.I.C, Nadia with the request to upload this on [www.nadia.nic.in](http://www.nadia.nic.in)
- 4) The Dy.CMOH-I/II, Nadia
- 5) The Accounts Officer, O/o the CMOH, Nadia
- 6) The IT-coordinator, Swasthya Bhawan with the request to upload this on [www.wbhealth.gov.in](http://www.wbhealth.gov.in)
- 7) The Senior Pathologist, District Public Health Laboratory, Shaktinagar Campus, District Hospital, Nadia. with the request to ensure successful delivery & installation of the bio-safety cabinet and issue Consignee Receipt Certificate and Satisfactory Installation Certificate
- 8) Notice-board of this office

  
**Chief Medical Officer of Health**  
Nadia S. Chanda