



Government of West Bengal
Office of the Principal
College of Medicine & Sagore Dutta Hospital
Kolkata – 700058; Phone: 033-25239058. Fax: 033-25239057
Website: www.cmsdh.edu.in; e-mail: principal.sdmch@gmail.com

Memo No. CMSDH/Pr/420/19

Date: 23-02-2019

CORRIGENDUM eTENDER NOTICE

Purchase and Installation of Complete Laparoscopic Set with all accessories including HD Camera & HD Monitor

NIT No:-36/2018-19

Tender ID No.: 2019_HFW_217188_1

Tender Reference No: CMSDH/Pr/383/19

Following modifications in Technical Specifications have been made

SI No.	Page No.	Exiting	Modification Done
01	9	Specification for full high definition Digital Camera	Specification for full high definition Camera- 3 Chip
02	10	The Camera Head Weight<100gms	Omitted the Camera Head Weight
03	10	Monitor Size not mentioned	Monitor size- Minimum 26 inches
04	11	LED Light Source- Minimum of 240 Watt	LED Light Source- Minimum of 300 Watt
05	11	Light Engine Type: Red, Green ,Blue LEDs(should have LED's for light emission)	Omitted this point
05	11	Electronic Scope Sensing Technology(ESST)	Omitted the Electronic Scope Sensing Technology(ESST)
06	11	Intuitive simple user interface with LCD touch screen	Intuitive simple user interface with LCD/LED touch screen
07	12	10mm 30 degree Precision eye-both 0 and 30 degree	Omitted both 0 and 30 degree(Read only 30 degree)
08	12	45 litter of high flow & having LCD Display measurement	45 litter of high flow & having LCD/LED Display measurement
09	13	Warranty-2year and CAMC-3year	Warranty-1 year and CAMC-3year(CAMC rate should be quoted separately as specified in NIT)

Except these, all other points and tender schedule will remain same.

HOD, Dept. Of Surgery,
College of Medicine & Sagore Dutta Hospital,
578, B T Road, Kolkata – 700058

Principal
College of Medicine & Sagore Dutta Hospital,
578, B T Road, Kolkata – 700058