



**Government of West Bengal**  
**Office of the Principal**  
**College of Medicine & Sagore Dutta Hospital**  
**578, B.T.Road, Kamarhati, Kolkata-700058**

Memo No: CMSDH/1350/21

Date: 09-07-21

**Corrigendum Tender Notice**

NIT No: CMSDH/04-2021-22

Tender Name: Supply, Installation and Commissioning of Automated Microbial Growth Detection System

Following points in technical specification are hereby modified:-

Sl No	Sl No in Specification	Existing	Modification
1	3	System can hold 30-40 samples at a time	System can hold <b>up to 120</b> samples at a time
2	5	The system should be able to detect fungal, aerobic and anaerobic organism from the blood	The system should be able to detect fungal, aerobic and anaerobic organism from the blood and <b>sterile body fluids</b>
3	13	System should be able to print reports on-monthly percentage of positive and negative culture, Monthly percentage of blood culture contamination, alert list and cumulative report of positive and negative culture.	Omitted
4	14	System should have a provision for blood volume monitoring.	Omitted



*DM 09/07/2021*

Principal

College of Medicine & Sagore Dutta Hospital  
Kamarhati, Kolkata-700058

**Principal**  
**College of Medicine**  
&  
**Sagore Dutta Hospital**  
Kamarhati, Kolkata-58