

Government Of West Bengal
Office of the Principal, Burdwan Medical College.
Purba Bardhaman.

Memo No. BMC/PC/

Dated, Purba Bardhaman, the

2023.

QUOTATION

Sealed quotation are hereby invited to the Dealers or Suppliers, for supply the undermentioned article for **Microbiology Department** at Burdwan Medical College, Purba Bardhaman. Rates should be inclusive of taxes, delivery charges and other charges if any, failing which the quotation will be treated as cancelled. Quotation should be reached this office on or before 29.05.23 within 1PM in a sealed cover mentioning the purpose of Quotation. No quotation will be received after the stipulated date and time. The undersigned reserves the right to accept or reject any quotation without assigning any reason. Documents to be produce viz 1 Valid trade license, 2.P.Tax & Pan Card (Xerox Copy)3. GST Registration, should be submitted with the rate quotation.

SI No.	Name of the Equipment	Required Quantity
1.	Elisa Reader (Specification attached herewith)	1 No.

Principal,
Burdwan Medical College.
Purba Bardhaman.

Memo No. BMC/PC/ 1342/1(7)

Dated, Purba Bardhaman, the 22/5/2023

Copy forwarded to:-


1. The District Magistrate, Purba Bardhaman.
2. The information and cultural officer, Purba Bardhaman is being requested to upload the Quotation notice in District NIC Website.
3. The Chief Medical Officer of Health, Purba Bardhaman.
4. The M.S.V.P., B.M.C. & hospital, Purba Bardhaman.
5. www.wbhealth.gov.in , Swasthya Bhawan, Salt Lake, Kolkata-700091
6. www.burmed.org , Purba Bardhaman.
7. College Notice board, for favour of information with the request to please display the notice in the Notice Board for wide circulation.

Principal,
Burdwan Medical College.
Burdwan

ELISA READER TECHNICAL SPECIFICATION

1. Display: Touch Screen & LCD not less than 5.6"
2. Operating System: Windows, User Friendly.
3. Absorbance Range: 0.0004-4.000 Abs
4. Reading Speed: <5 seconds for 96 well plates (single wavelengths)
5. Storage Capacity: More than 10000 Records.
6. Warm up time: No warm up time
7. Printer: In built Thermal Printer and External Printer optional.
8. Input: Touch screen with stylish pen & mouse.
9. LIS & HIS Compatibility: Compatible with application support.
10. Data Protection (Auto back up in case of power outage): Available
11. Lamp Sleeping & Wake up function: Available
12. 8 Chanel optical fiber system: Yes
13. Test Results: Result will be shown with measuring range set for the individual test parameter.
14. Plate Carrier: Covered with black lid to avoid interference of external light.
15. Manual entry of the O.Ds possible: Yes
16. Wavelength & no of filter: 405, 450, 492, 630 with 3 optional filters.
17. Calculation Mode: Absorbance, Cut Off, Single Point Calibration, Curve Regression, Index Regression, Multiple Point Percentage, Linear Regression, Log Regression, Percentage Regression, Four Parameter Regression & Logistics
18. Multi Assay: Upto 12 different assay on 1 Plate
19. Light Source Provided: Tungsten Halogen
20. Type of Model: Bench Top
21. Read Method: End Point & Kinetic
22. Blanking Facility Provided: Well Wise
23. Internal Calibration: 9 Point Calibration
24. Wells Reading Capacity: U, V & Flat typewells
25. Warranty: One year
26. Certification: CE-IVD


Principal
Burdwan Medical College
Purba Bardhaman


Head of the Department
Department of Microbiology
Burdwan Medical College