



Government of West Bengal
Office of the Principal
Medical College, Kolkata – 700 073

Memo No. MC/Prin/ 2825 /2022

Dated: 23/05 /2022

TENDER NOTICE

Principal, Medical College, Kolkata, invites Sealed Tender from reputed Suppliers to supply of **One Autoclave Machine (Semi-automated)** for the Dept. of M.R.U. at Medical College, Kolkata for Research Project. Tender details including specification can be obtained from the Departmental website www.wbhealth.gov.in from 23/5/2022. Last date of submission of Tender is 30/6/2022 up to 2.00 p.m. Sealed Tender will be opened on the same date at 2.30 p.m. at the chamber of Principal, Medical College, Kolkata in presence of the intending bidders.

Rm 23/05/2022
Principal

P Medical College, Kolkata

Memo No. MC/ Prin/ 2825 /2022/1(2)

Dated: 23/05 /2022

Copy forwarded for information and necessary action to:-

1. The Accounts Officer, Medical College, Kolkata.
2. I.T.Cell, Swasthya Bhavan, Salt Lake, Kol-91, for publication of tender in the departmental website

Rm 23/05/2022
Principal

P Medical College, Kolkata

Terms & Conditions

Tender Memo No: MC/Prin/ 2825 Dt. 23/05/2022

1. Manufacturers, Authorized dealers, Distributors, Firms are eligible for quoting in this tender as per specification of the equipment attached here.
2. The rate should be quoted in both figures & words. Rates quoted should be all inclusive of delivery charges to the concerned department including installation & packing charges etc. All applicable charges should be mentioned separately including GST.
3. Earnest Money amounting Rs.10000/- (Ten thousand) only must be deposited in the form of DD from any Nationalized Bank in favor of the Principal, Medical College payable at Kolkata by the intended bidders.
4. The Bidders must submit their tender in two bid method. The main envelope should contain Technical Bid in envelop A and Financial Bid in envelop B. All envelopes must be properly sealed.
5. Price Bid of those bidders will be opened only after qualifying in technical scrutiny & physical demonstration according to the decision of tender selection committee.
6. Successful bidder will deposit _____ % amount of the articles as a security deposit from the date of supply.
7. The tenderer shall be bound for total replacement of the supplied quantity with a fresh supply of equivalent quantity if the item(s) is/are found to be defective/not as per given specification.
8. Specification & brochure is to be submitted along with the tender.
9. The warranty of the equipment will be at least two years. The rate of AMC/CMC of equipment of three years (from 3rd yr. to 5th yr. after warranty) has to be furnished separately without submitting the tender will be treated invalid.
10. The Tendering Authority reserves the right to accept /reject any or all tender without assigning any reason whatsoever.
11. Payment shall be made only after the delivery of the articles and satisfactory installation certification given by the concerned departmental authority.
12. All essential documents as stated below have to be submitted by the bidder along with technical bid. Failing to submit these documents the bid will be treated as cancelled.
13. The last date of submission of sealed tender is _____ 2022 up to 2.00 pm in the tender box kept in the chamber of Accounts Officer, Medical College. Tender will be opened on same date at 2.30 P.M. at the chamber of the u/s in presence of the intending bidder.
14. The specification of equipment is being attached herewith the tender.
15. The lowest rate will be accepted.

List of essential documents to be submitted:

- Copy of GST registration certificate showing registration no.
- Copy of valid current trade license
- Copy of pan card of the bidder
- Copies of I.T. return for last two years
- Authorization letter from Manufacturer, if any
- Credential of the bidder in successful supply to any Govt. hospital

Ph 23/05/2022
Principal
Ph Medical College, Kolkata

Technical specification for Autoclave (Semi-automated)

Specifications:

- Double wall construction
- With alarm for low water level
- Automatic purging for efficient sterilization
- Larger size with convenient raise for lid lifting
- With SS wire mesh carrier and spare gasket
- Maximum operating pressure meter 103.4 kPA
- Operating volume 110 liters or more
- It should be CE certified
- Warranty: 1 year

Sw

18/5/22

Biswabandhu Barik
14/5/22

m
18/5/22

Member Secretary
MRU, Medical College, Kolkata