

GOVERNMENT OF WEST BENGAL
OFFICE OF THE MEDICAL SUPERINTENDENT CUM VICE-PRINCIPAL
R.G.Kar Medical College & Hospital, Kolkata

Memo No.: RGKH/ 3104

Dated 23.5.22

NOTICE INVITING QUOTATION

Sealed "Quotation" are invited from the reputed Company/firm/ authorized distributor firms for the supply of Electronic Balance ,Single Pan for Biochemistry Department of R.G.Kar Medical college and Hospital, Kolkata.

The last date of submission of quotation is 30.5.2022 on 1 PM and the same will be opened on the same date at 2.00 PM in presence of interested bidder, if any. The bidders will submit the Quotation in their own official pad along with the following essential documents :


1. Valid Trade License/Enlistment.
2. PAN card of the Bidder/Bidder Company.
3. GST Registration certificate.
4. 4..I.T Return of Last year .
5. Professional Tax paid Challan

Specification in detail:-1) Electronic Balance is required for precision weighing of lab sample.2.Microprocessor based single pan Analytical Balance with high accuracy & precision is required.3.Reading the weight by digital display.4.Electronic Top Loading balance with transparent case,.5.The Balance should have function of piece counting ,percent weighing formulation ,dynamic weighing with automatic and manual start and provision for data interface.6. weighing accurately upto 4th decimal place. 7. Auto self calibration facility 8. Auto Zero setting, 9. One touch calibration.10. Weighing capacity upto 200gm.10.Reapitability and resolution :0.1mg, 11. Linearity:+0.2mg 12.Stabilisation time<5sec.13Adjustment weight (Int. Wt) :-200g. 14. Adjustment weight.(Ex. Wt) :-500gm, 1gm,10gm,50gm,100gm,200gm. 15. Balance should have the following feature:-LCD 7 segment Backlit Display , Should be calibrated with external wt. Within a frequently use partial range., Vibration adapter for damp influence due to vibration & minor shocks, Chemical resistant housing, 16. Should be supplied with standard external & internal wt. As specified shall meet IEC-60601-1-2:2001/ BIS.,16. Power Input 220-240 V Ac 50 Hz Fitted with Indian Plug, UPS of suitable rating with voltage regulation and spike protection for 60 minute back up.17. Should be comply with ISO/GLP with auto validation with ink jet printer. Should be FDA, CE, UL, or BIS Approved product.18.List of Important spare and their part no. And costing.

The quotation is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 23.5.2022


The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason. Rate should be quoted in rupees only.

Cumulative Lowest rate will be accepted.


Medical Superintendent-cum-Vice Principal
R.G.Kar Medical College & Hospital, Kolkata
Dated 23.5.22

3104
Memo No.: RGKH/ 3104
Deputy Superintendent
R.G. Kar. MC&H
Kol-4

- Copy to:-
1. Principal, R.G.Kar, M.C.H, Kolkata -700004
 2. Deputy Superintendent, R.G.Kar, M.C.H, Kolkata -700004
 3. Account Officer, R.G.Kar, M.C.H, Kolkata -700004
 4. Notice Board, M.S.V.P. Office, R.G.Kar, M.C.H, Kolkata -700004
 5. System Co-ordinator . I.T. Cell ,Swasthya Bhawan ,Salt Lake, with request to upload this notice in the departmental website .


Medical Superintendent-cum-Vice Principal
R.G.Kar Medical College & Hospital, Kolkata

Memo No.: RGKH/

Principal, R.G.Kar Medical College & Hospital, Kolkata -700004

Deputy Superintendent, R.G.Kar Medical College & Hospital, Kolkata -700004