



Govt. of West Bengal
Office of the Principal
College of Medicine & Sagore Dutta Hospital
Kamarhati, Kolkata-700058

Ph no: 033-2583-4277

Memo No:CMSDH/Pr/674/19

Date: 12-03-2019

Quotation Notice

The sealed quotation is invited from the reputed organization for supply & Installation of following Item for Central Pathology and Bio-chemistry department under College of Medicine & Sagore Dutta Hospital, The details of above item is as follows-

Name of the Item	Specification	Quantity
1. Auto pipette(variable volume) (Autoclavable)	as per annexure- I	as per requirement.

Type:

- 100ul- 1000ul.
- 5ul-50ul
- 10ul-100ul

The rate should be inclusive all taxes and others charges (Installation) if any.
The following documents should be submitted along with quotation Letter.

1. Xerox copy of valid Trade license in related field.
2. Xerox copy of GST Registration
3. Xerox copy of PAN card
4. Last year IT Return

The quotation must reach this office (Box kept at Principal's office Room) on or before 18.03.2019, latest by 2.00 pm with super scribed on the top of the envelop . Quotation will be opened on 18.03.2019 at 3.00 PM

- If Opening date be holiday then quotation/tender will be open on next working day at same time

The undersigned reserves the right to accept or reject any quotation without assigning any reason.

The quotation notice is to be Downloaded from college website- (www.cmsdh.edu.in)& (www.wbhealth.gov.in).

Principal

College of Medicine & Sagore Dutta Hospital
Kamarhati, Kolkata-700058

Memo No: - CMSDH/Pr/674/19/1(6)

Date: 12-03-2019

Copy forwarded for necessary information:-

1. MSVP, College of Medicine & Sagore Dutta Hospital.
2. CMOH, North 24 Parganas
3. Accounts Officer, College of Medicine & Sagore Dutta Hospital
4. Assistant Superintendent, College of Medicine & Sagore Dutta Hospital
5. Notice Board, College of Medicine & Sagore Dutta Hospital
6. The IT Cell, Swastha Bavan, Saltlake, kol-91, with request to upload the quotation notice in departmental website.

12/3/19

Principal

**College of Medicine & Sagore Dutta Hospital
Kamarhati, Kolkata-700058**

Annexure I
Specification of Variable Volume Auto Pipette
(Autoclavable) for Bio-chemistry

1. Robust and easy to use
2. Soft good grip (material:TPE)
3. High chemical resistance
4. Fully autoclavable
5. Volume setting by click stops
6. Suitable for both left and right-handed
7. Compatible with a wide range of tips
8. Volume range color-coding for easy pipette and tip selection
9. Manufactured as per ISO 9001:2008 & ISO 13485. Each pipette should be individual
10. Calibration certificate (NABL) must be provided with each pipette.
11. CE marked
12. 3-years warranty

- Volume range: Variable Volume micropipette: 5-50 μ l
Increments (μ l): 0.1 μ l
Accuracy (\pm) %:0.80 CV (\pm) %:0.4

- Volume range: Variable Volume micropipette: 10-100 μ l
Increments (μ l): 0.2 μ l
Accuracy (\pm) %:0.60 CV (\pm) %:0.2

- Volume range: Variable Volume micropipette: 100-1000 μ l
Increments (μ l): 1.0 μ l
Accuracy (\pm) %:0.60 CV (\pm) %:0.2