

Government of West Bengal
Office of the Principal
NRS Medical College, Kolkata

Memo No: NMC/Store/ 696

Date: 13 / 02 / 2018

Tender Notice
(NIT NO. 23/2017-18)

Sealed Tender is invited by the Principal, N. R. S. Medical College, Kolkata – 14 for supply of **ULTRASOUND THERAPY UNIT**. Tender form and details/specification can be obtained from the Departmental website www.wbhealth.gov.in from 13/02/2018. Last date of submission of Tender is 20/02/2018 up to 1 P.M. Tender will be opened on same date at 2.30 P.M in the office of the Principal, N. R. S. Medical College. Details are available in the Department of Health & Family Welfare, Govt. of West Bengal official Website: www.wbhealth.gov.in & College website (www.nrsmc.edu.in)




Principal
NRS Medical College,
Kolkata-14

Memo No: NMC/Store/ 696/2(4)

Dated: 13/02/2018


1. The Medical Superintendent-cum-Vice Principal, NRSMCH, Kolkata
2. Accounts Officer, NRS Medical College, Kolkata
3. I. T. Cell, Swasthya Bhawan, Salt-lake, for publication in the website.
4. Notice Board.



Principal
NRS Medical College
Kolkata-14


Terms and Condition

1. The rate should be quoted in both figure and word strictly as per sample/mentioned specification along with provided accounting unit only. All rates quoted should be inclusive of all livable charges (including door delivery charges to the concerned department, packing charges, installation charges), and all other foreseeable incidental charges are exclusive of all applicable taxes should be mentioned separately.
 2. The Tendering authority reserves the right to withdraw any item from the Tender at any stage. In such Contingency the selection of such item, if already made in favor of any Tenderer shall be treated as cancelled. The Tendering authority also reserves the right to accept / reject / split any.
 3. The Tenderer shall be bound for total replacement of the supplied quantity with a fresh supply of equivalent quantity if the item(s) are found to be defective/not as per given specification.
 4. Payment will be made only after the delivery of the articles and satisfactory certification by an authority fixed by the administration. No request for advance or part payment will be considered.
 5. In case of rejection of any supply on quality grounds or failure or refusal to replace the contracted items within the specified time or according to specifications/quality, the College/Hospital administration will have the authority to recover the losses, if any, by making deductions from any pending claim of the supplier in respect of same tender or any past/future tender.
 6. The Specification and Brochure is to be submitted along with the Tender.
 7. The Tenderers must accompany all the papers/documents along with the Tender.
- The Tenderers must submit their tender in two bid method. Main envelope should contain Technical bid in Envelop A and Financial bid in Envelop B. all the envelopes must be properly sealed before submission.
 - Price bid of those bidders will only be opened who will qualify in technical scrutiny by tender selection committee.
 - Earnest money amounting to Rs. 5,000/- (Five thousand) only must be deposited in the form of Demand Draft from any Nationalized Bank in favour of 'Principal, N. R. S. Medical College' payable at Kolkata by L1 bidder before collection order.


Principal
NRS Medical College
Kolkata-14

List of essential documents to be submitted

- Copy of GST Registration certificate showing Registration No.
- Copy of Valid Trade license
- Copy of PAN Card of the bidder
- Copies of Income Tax return for the last three years (Assessment year: 14-15, 15-16,16-17)
- Copies of Profit & Loss A/c and Balance Sheet for the last three years (2014-2015, 2015-2016, 2016-2017)
- Credential of the bidder in successful supply to Govt. hospitals
- Bank Solvency Certificate



Principal
NRS Medical College
Kolkata-14

Instruments required for Hemophilia patient care.

Page 1, Page 2

1. Ultrasound Therapy Unit

Specifications:

- Output Mode: Continuous & pulsed
- Pulsed Mode: 1:1, 1:2, 1:4 and so on.....
- Transducer: Separate Transducer for 1 MHz & 3 MHz
- Parameter selection: Manual & Programmed
- Intensity: Linear adjustment 1- 3.5 watt/Cm²
- Patient safety: Output is 100% isolation
- Pre program: 30 pre programmed for 1 MHz 10, for 3 MHz 20
- Output display: Back Light LCD
- Treatment Time: Digital time pre settable 1-20
- Output display: LCD Display
- Power Supply: standard

Special Features

- Compact Size,
 - Provision for combination therapy facility.
 - Onsite warranty and AMC for 5 years.
-