

Bid Document

Bid Details	
Bid End Date/Time	30-03-2022 17:00:00
Bid Opening Date/Time	30-03-2022 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	60 (Days)
Ministry/State Name	West Bengal
Department Name	Health And Family Welfare Department West Bengal
Organisation Name	General Administration
Office Name	Swasthya Bhawan
Total Quantity	134
Item Category	Clinical incubators or infant warmers (Q3)
Minimum Average Annual Turnover of the Bidder	41 Lakh (s)
OEM Average Turnover (Last 3 Years)	329 Lakh (s)
Years of Past Experience required	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	30 %
Bid to RA enabled	No
Time allowed for Technical Clarifications during technical evaluation	3 Days
Evaluation Method	Total value wise evaluation

EMD Detail

Advisory Bank	ICICI
EMD Percentage(%)	2.00
EMD Amount	164820

ePBG Detail

Advisory Bank	PUNJAB NATIONAL BANK
ePBG Percentage(%)	3.00
Duration of ePBG required (Months).	30

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

PO-II, NHM & Deputy Secretary, H&FW Samity
Swasthya Sathi Building, Swasthya Bhawan, Health and Family Welfare Department West Bengal, General Administration, GN-29 Block, Sector-V, Kolkata-700091
(Brijit Suchita Kujur)

Splitting

Bid splitting not applied.

MSE Purchase Preference

MSE Purchase Preference	No
-------------------------	----

MII Purchase Preference

MII Purchase Preference	No
-------------------------	----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant

contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Clinical Incubators Or Infant Warmers (134 pieces)

Brand Type	Unregistered Brand
------------	--------------------

Technical Specifications

* [As per GeM Category Specification](#)

Specification	Specification Name	Bid Requirement (Allowed Values)
Technical Characteristics (specific to this type of device)	Type of heater:	quartz infrared
	Audiovisual alarm facility for:	Overheating beyond set temperature range, Spurt in Heater output i.e. more than 60% heater output for 10 minutes system should raise alarm and cutoff heater, Power failure, Heater failure, Patient temperature less/Greater than required/set temperature, Probe failure
	Warmer head should be rotatable in different direction, so as to allow taking X-ray	Yes
ENERGY SOURCE	Compatible to alternate energy supply source; Solar Heating in addition to AC power source):	Yes, No
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	3, 5, 2, 4 Or higher

Additional Specification Parameters - Clinical Incubators Or Infant Warmers (134 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Heater with Microprocessor based servo controlled unit ,LED display , Rechargeable Battery backup for display of temperatures & alarm during power failure,Adjustable light LED source with manual angulation (for observing baby & IV cannulation)	2 Operation modes – servo (automatic) and manual mode, 1” display for showing skin temp in manual mode and set and skin temp in servo mode,Temp. display- bright, numerical, Heater output control: 0 -100% in increments of 10% or lower in manual mode.The unit should stop heating if the temp exceeds the desired value by 1deg and restart only when temp falls back below set temp.

Specification Parameter Name	Bid Requirement (Allowed Values)
Bassinets with Stitchless Mattress (made of fire retardant material & foam density of 21-25 kg/m ³ with 1" thickness.) & 1 SS-304 grade IV stand (max dia of IV stand (SS) - 2.5 cm) for hanging IV fluids, holding 03 Syringe pump	For easy swivel movement, separate from warmer & movable on lockable wheels, 1 fixed transparent acrylic sheet below the mattress for under surface phototherapy & an additional acrylic tray with the bassinet for positioning the X-ray cassette, Inclination in both directions (trendelenburg and reverse trendelenburg) min upto 12°, Mattress: waterproof & easily washable & resistant to chlorine based solution, 2 drawers and 4 pegs (SS) of IV Stand with power source with top (5 amps & 15 amps, 4 nos.)
Physical Parameters	Probe Length > 5.5 ft., Probe sense Dia Max. 1 cm, Bassinet (LXW): Min (680 mm X 500 mm, Working level (ground to upper level of the mattress): Min 870 mm, Distance b/w mattress top & heat source: Min 850 mm, Height adjustment system upto ± 2 inch, Dia of castors with brakes / locks on front wheels: Min 75 mm, Monitor tray: 300 mm X 270 mm, Upper Drawer: 400 mm X 430 mm X 120 mm, Lower Drawer: 400 mm X 430 mm X 145 mm, Foldable transparent acrylic side panels - min 6 mm thickness
Certification	FDA or CE certificate, Test certificate & ISO 13485 certificate.
Accessories & Consumables to be supplied with the machine (Price for 1 additional monitor tray should be given separately (including all accessories and consumables))	Mattress: 02, Reusable Temp. Probe: 02, Fuse set: 01, Monitor tray: 01

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Debashree Ghosh	700091, Swasthyasathi Building, Swasthya bhawan, GN-29, Sector-V, Salt lake, kolkata	134	45

Buyer Added Bid Specific Terms and Conditions

1. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specification and / or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents / clauses

shall also be null and void. If any seller has any objection / grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---