

Bid Document

Bid Details	
Bid End Date/Time	30-03-2022 17:00:00
Bid Opening Date/Time	30-03-2022 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	60 (Days)
Ministry/State Name	West Bengal
Department Name	Health And Family Welfare Department West Bengal
Organisation Name	General Administration
Office Name	Swasthya Bhawan
Total Quantity	77
Item Category	Phototherapy machine for SNCU (Q3)
Minimum Average Annual Turnover of the Bidder	15 Lakh (s)
OEM Average Turnover (Last 3 Years)	121 Lakh (s)
Years of Past Experience required	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	30 %
Bid to RA enabled	No
Time allowed for Technical Clarifications during technical evaluation	3 Days
Evaluation Method	Total value wise evaluation

EMD Detail

Advisory Bank	ICICI
EMD Percentage(%)	2.00
EMD Amount	60984

ePBG Detail

Advisory Bank	PUNJAB NATIONAL BANK
ePBG Percentage(%)	3.00
Duration of ePBG required (Months).	30

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

PO-II, NHM & Deputy Secretary, H&FW Samity
Swasthya Sathi Building, Swasthya Bhawan, Health and Family Welfare Department West Bengal, General Administration, GN-29 Block, Sector-V, Kolkata-700091
(Brijit Suchita Kujur)

Splitting

Bid splitting not applied.

MSE Purchase Preference

MSE Purchase Preference	No
-------------------------	----

MII Purchase Preference

MII Purchase Preference	No
-------------------------	----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant

contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Phototherapy Machine For SNCU (77 pieces)

Brand Type	Unregistered Brand
------------	--------------------

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Bid Requirement (Allowed Values)
Performance Parameters	Type of phototherapy machine	On Mobile Floor stand
	Type of coolinf	Natural cooling, Forced air circulation
	The range of main radiation spectrum in nm	400-550, 420-500
	Minimum spectral Irradiance in $\mu\text{W}/\text{cm}^2/\text{nm}$	35 - 40 $\mu\text{W}/\text{cm}^2/\text{nm}$ at 45 cm height
	Availability of digital hour meter showing time	Patient Treatment time, Treatment time for patient + Cumulative lamp hours
	Average Effective light field in mm^2	700-1000, 1000-1500, 1500-2000
	LED life hours for LED light unit in hours	above 20000-upto 30000, above 30000-upto 40000
	Facility of Over temperature safety cut out	Yes
	The irradiance ratio (min to max) on mattress	30 - 40 %, more than 40%
	Free fall protection for LED	Spring Counter balance with locks, Counter balanced with locks
	Facility of adjustment of light intensity	Variable in steps
	Number of castors	4
	Number of brakes	2
	Phototherapy unit shall be equipped with white light for observation purpose	Yes
	Phototherapy unit shall be provided with irradiance measurement meter in $\mu\text{W}/\text{cm}^2/\text{nm}$	With external flux meter, With inbuilt flux meter, Not provided
Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	2, 3, 4, 5, 1	

Specification	Specification Name	Bid Requirement (Allowed Values)
Additional Parameters	Power consumption in watts	upto 45, above 45 - upto 60, above 60 - upto 100
	Number of replacement sets of fuses	1, 2, 4, Not provided
	Number of Infant eye masks of different sizes	250, 500, 750, 1000, Not provided

Additional Specification Parameters - Phototherapy Machine For SNCU (77 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Design with Under surface Phototherapy Facility	i)Sturdy mobile stand with antistatic castors ii)On/Off switch to start/stop phototherapy iii)Tilting facility both horizontally and vertically at any angle iv)Height adjustability- Min 138 cm and max 190 cm (\pm 5%)& made of SS v) It must have a Sticker, in front of the phototherapy unit as desired,approved from NHM before supply vi)The base stand should occupy less space vii)LED life: >50,000hrs viii)Power Consumption: <30 Watts
Irradiance	i)Should be adequate and uniform across the bassinet i.e. clinically required irradiance should cover the full length of the baby from head to toe ii) The supplier will measure irradiance and show the irradiance to end user each time during his quarterly visit using a standard calibrated irradiance meter which measures 450 – 480 nm wavelengths. The irradiance is to be measured at a distance of 45 cm from the lamp.
Spares & Accessories	Should supply reusable Bubble Sheet - 04 nos. with the equipment
Certification	Complete unit should conform to internationally accepted quality standards and should carry the US FDA or CE (“Conformite Europeene”) from European Union notified body having 4 digit identification number certification.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Installation Commissioning and Testing (ICT) details for the above item:

% of Product Cost Payable on Product Delivery	80%
Min Cost Allocation for ICT as a % of product cost	20%
Number of days allowed for ICT after site readiness communication to seller	45 Days

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Debashree Ghosh	700091,Swasthyasathi Building,Swasthya bhawan, GN-29, Sector-V, Salt lake, kolkata	77	45

Buyer Added Bid Specific Terms and Conditions

1. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specification and / or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents / clauses shall also be null and void. If any seller has any objection / grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions.](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---