

Bid Document

Bid Details	
Bid End Date/Time	30-03-2022 17:00:00
Bid Opening Date/Time	30-03-2022 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	60 (Days)
Ministry/State Name	West Bengal
Department Name	Health And Family Welfare Department West Bengal
Organisation Name	General Administration
Office Name	Swasthya Bhawan
Total Quantity	187
Item Category	Pulse Oximeters (Q3)
Minimum Average Annual Turnover of the Bidder	205 Lakh (s)
OEM Average Turnover (Last 3 Years)	1645 Lakh (s)
Years of Past Experience required	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	30 %
Bid to RA enabled	No
Time allowed for Technical Clarifications during technical evaluation	3 Days
Evaluation Method	Total value wise evaluation

EMD Detail

Advisory Bank	ICICI
EMD Percentage(%)	2.00
EMD Amount	822800

ePBG Detail

Advisory Bank	PUNJAB NATIONAL BANK
ePBG Percentage(%)	3.00
Duration of ePBG required (Months).	30

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

PO-II, NHM & Deputy Secretary, H&FW Samity
Swasthya Sathi Building, Swasthya Bhawan, Health and Family Welfare Department West Bengal, General Administration, GN-29 Block, Sector-V, Kolkata-700091
(Brijit Suchita Kujur)

Splitting

Bid splitting not applied.

MSE Purchase Preference

MSE Purchase Preference	No
-------------------------	----

MII Purchase Preference

MII Purchase Preference	No
-------------------------	----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant

contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Pulse Oximeters (187 pieces)

Brand Type	Registered Brand
------------	------------------

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Bid Requirement (Allowed Values)
Performance Parameters	Type of Pulse Oximeter	TABLE TOP
	Special Signal processing Technology (Motion artifact rejection)	Masimo - SET technology
	SPO2 probes	Reusable
	Display of SPO2 and PR	digital display with plethysmograph, digital display with plethysmograph with trends
	Type of Patient	Adult, Pediatric, Neonatal
	Should be shock proof	Yes
	Display (with contrast adjust-ability)	LCD
	Display should be Readable from at least 3 feet	Yes
	Saturation displayed on the screen	Yes
	Pulse Rate displayed on the screen	Yes
	Perfusion Index displayed on the screen	Yes, No
	Status of battery charging displayed on the screen	Yes
	Sensor off /on status	Yes
	Audible and visual Alarms	Yes
	Alarm override facility should be present	Yes
	Trends	48 hrs, 96 hrs
	Type of trend display	Tabular, Graphical
Power source	Inbuilt re-chargeable battery and mains 220-240 V , 50 Hz AC operation	
Automatic switch from mains to battery in case of power failure	Yes, NA	

Specification	Specification Name	Bid Requirement (Allowed Values)
	RS 232C interface for data communication and transfer	Yes, NA
	Should have provisions for wireless and blue tooth connectivity	Yes, NA
	Patient extension cable	Yes
	Protective splash proof case for clean and storage and safe transport provided	Yes, No
	Length of power cable	3m or more
Probes	Type of probe	Neonatal
	Probe usage	Reusable
	Number of probes provided (Hint: in case adult , pediatric and neonatal are selected for type of probe all the three have to be given and in case adult & pediatric are selected both are to be given)	1 each, 2 each, 3 each, 4 each
NIBP Parameters	Non invasive blood pressure (NIBP) provided	Yes, No
	Type of BP monitoring	Adult, Pediatric, Neonatal, NA (If NIBP not provided)
	NIBP range in mmH	20-400, NA (If NIBP not provided)
	Cuff size	Adult, Pediatric, Neonatal, Extra Large, NA (If NIBP not provided)
	Number of cuff's provided	1 each, 2 each, 3 each, 4 each, NA (If NIBP not provided)
	Warranty in years (without probes)	2, 5
	Warranty for probes in years	1, 2, Not Provided
Reports	Confirmity to Manufactures Certification	ISO 9001 & ISO 13485

Additional Specification Parameters - Pulse Oximeters (187 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)

Specification Parameter Name	Bid Requirement (Allowed Values)
Design & Display size	1. Portable, light weight and desktop pulse oxymeter. 2.A large LED / LCD display (min 4 ") so that numerical values are readable from a distance of at least 10 feet with naked eyes 3. SpO2 range - 1 to 100%. 4. Visual alarm for both upper and lower SpO2 and Pulse rate and adjustable audio alarm volume. 5.Can capture data for real time analysis at least for 24 hrs. 6. Patient type: Neonatal only
Accessories	<ul style="list-style-type: none"> Should supply Ambient Light cover for sensor - 20 nos. with each machine The following SpO2 probes and cable should also be supplied with each machine: a) a) Reusable neonatal probes (flex with moulded or detachable type rubber or silicon made): 2 (two) nos. b)Finger probe including cable: 1 (one) nos. The cable length should not be less than 3 meters.
Data Analysis	i)Signal delay(Response time): <15sec ii)Averaging time: <3sec respectively iii) Accuracy(for Spo2 & Pulse rate): +_3digits iv) Frequency of capturing data(trends & future review): Every 4 sec interval
Standards,Safety and Training	i. USFDA & CE certificate ii. ISO certificate iii.Should have onsite service facility. The service provider should have the necessary equipment recommended by the manufacturer to carryout preventive maintenance test as per guidelines provide in the service / maintenance manual. iv. Environmental Factors Operating temperature : 10°C to 40°C Storage temperature : -10°C to 50°C Operating humidity: 15% to 95%
Warranty	Warranty for 2 years and 5 years of CMC after warranty. The warranty and CMC includes entire equipment including battery, SpO2 extension cable etc. except probes.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Debashree Ghosh	700091,Swasthyasathi Building,Swasthya bhawan, GN-29, Sector-V, Salt lake, kolkata	187	45

Buyer Added Bid Specific Terms and Conditions

1. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specification and / or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline

governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents / clauses shall also be null and void. If any seller has any objection / grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---