

Govt. of West Bengal
Office of the Superintendent, Vidyasagar S. G. Hospital
Behala, Kolkata – 700034.

Memo No. VHB/ 904

Dated: 19.08.2020

TENDER NOTICE

Sealed tenders are invited by Superintendent, Vidyasagar S.G. Hospital on behalf of Govt. of West Bengal, from bonafide agencies for supply of Multi Para Monitor (ECG/SPO2/NIBP/RESP/TEMP) for Vidyasagar S.G. Hospital.

Quantity to be supplied: 2(two)

Forms along with terms and conditions and schedule documents will be available in the website <https://www.wbhealth.gov.in/vidyasagarsgh/> on and from to 20.08.2020 and the same (duly filled up) will be received either through speed post or in drop box except Govt. holidays in between 10 a.m. - 4 p.m. within 05.09.2020 upto 12.30 p.m. and the technical bids will be opened at 2 p.m. on the same day (05.09.2020). The financial bid will be opened at 2:30pm on 05.09.2020 Applications in due format will be received from bidders addressing the “**Superintendent, Vidyasagar S. G. Hospital, South 24 Parganas**”.

The amount of security deposit of 5% of the total value of purchase to be deposited by the successful bidder through demand draft in the name of “**Vidyasagar State General Hospital, Rogi Kalyan Samiti**” to the undersigned good office or this will be deducted from the bill at the time of payment and will be kept with the office of the undersigned till 06(six) month or the maturation of the next tender whichever is earlier. The rate of the tender will be final for the same product for one year or till the finalization of next EOI (Empanelment) from the date of acceptance of rate by the Tender Selection Committee whichever is earlier. The decision of the tender selection committee is final for acceptance or rejection of any tender without assigning any reason.

The technical and financial bid should be submitted separately.



Superintendent
Vidyasagar S. G. Hospital
Behala, Kolkata – 34

Memo No: VBH/

Date:

Copy forwarded for information to:

1. The C.M.O.H., South 24 Parganas
2. The SDO , Alipore Sadar
3. The Chairman, Municipality, Borough XIV
4. Notice Board , Vidyasagar S.G. Hospital
5. Office Copy


Superintendent
Vidyasagar S. G. Hospital
Behala, Kolkata – 34

Govt. of West Bengal
Office of the Superintendent, Vidyasagar S. G. Hospital
Behala, Kolkata – 700034.

Terms and conditions for procurement of 2 (two) no. of Multi Para Monitor for Vidyasagar S. G. Hospital. vide Memo No.VHB / 904 Dated. 19.08.2020

In case of necessity, the date of opening may be deferred, in that case notification will be displayed in the Notice Board in the Office of the undersigned in due time or in <https://www.wbhealth.gov.in/vidyasagarsgh/>

1. Name, Address and NIT No and purpose must be clearly mentioned on the sealed envelope.
2. Tenderer has to submit the sealed tender form at their own cost to the office of the undersigned within stipulated time through speed post only or in drop box and will be opened by the purchase committee in presence of the tenderer or his representative.
3. The claimed rate should cover transport and delivery charges. The tender must indicate whether the price quoted is inclusive of GST or not. Quoted rates should be clearly mentioned in both figures and words.
4. Cooperative society should furnish along with their tender current certificate from ARCS of concerned district regarding performance.
5. All sheets should be numbered properly and number of sheets in each cover should be cited in covering tally.
6. Pan No., GST No., Trade License is to be furnished with the tender documents.
7. The equipments supplied must comply all the specifications mentioned in ANNEXURE I & II, and must be a branded goods.
8. The Authorization of the Manufacturers is desired to be submitted by the bidders.
9. The amount of security deposit of 5% of the total value of purchase to be deposited by the successful bidder through demand draft in the name of “**Vidyasagar State General Hospital, Rogi Kalyan Samiti**” to the undersigned good office or this will be deducted from the bill at the time of payment and will be kept with the office of the undersigned till the maturation of the next tender or for six months whichever is earlier. The rate of the tender will be final for the same product for one year from the date of finalization of the rate by the Tender Selection Committee.
10. Technical bid and financial bid should be submitted in two separate envelopes mentioning the respective bids on the sealed envelopes.
11. Financial Bid of Technically fit Bidder will be opened only.
12. In case of deviation from any of the above terms and conditions, contract will be liable to be terminated by the appropriate authority.
13. The selection Committee reserves the right to accept/ cancel any tender without assigning any reason thereof.
14. Please refer to website for further corrigendum (if any) during the period of processing.


Superintendent
Vidyasagar S. G. Hospital
Behala, Kolkata – 34.

(SIGNATURE WITH STAMP)

19/8/20
Superintendent
Vidyasagar S. G. Hospital
Behala, Kolkata-34

ANNEXURE I

Signature of Issuing Authority with stamp

.....

.....

Application Form and technical documents (to be submitted with the technical bids)

Application Form :

Application form for bidders for supplying of 2(two) no. of Multi Para Monitor for Vidyasagar S. G. Hospital **VIDE NIT NO. VHB/** **Dated**

To
The Superintendent
Vidyasagar S.G. Hospital
Behala, Kolkata-700034.

Sir,

Reference: Your advertisement no _____ dated _____ in the _____ inviting tenders for Vidyasagar S.G.Hospitals.

With reference to the above, I/We wish to offer our tender for **supplying** of 2(two) no. of Multi Para Monitor for Vidyasagar S. G. Hospital. I/We have read all bid documents carefully and hereby declares that I/ we have the due expertise and financial capability to successfully undertake the contract if awarded. The bids are made in two parts- Technical and Financial and are separately enclosed. I/We agree to all the terms and conditions of the tender. I am also aware that if I indulge in any unfair practice or submit any fraudulent documents, my present bid will be cancelled and I will be debarred from participation in any future tenders for five years. Thanking you.

Yours sincerely,

ANNEXURE II

Financial Bid

la sm
Superintendent
Vidyasagar S. G. Hospital
Behala, Kolkata-34

To
The Superintendent,
Vidyasagar S.G. Hospital
Behala, Kolkata-34

Respected Sir,
I/We are submitting the best rate for Multi Para Monitor as per specifications mentioned below:

Specifications should match the following:-

The equipments to be supplied must be branded.
Quantity to be supplied2 (Two)

Price Bid for Multi Para Monitor (each quantity) for Vidyasagar S. G. Hospital.

Specification: PATIENT MONITORING SYSTEM FOR ADULTS, PEDIATRICS AND NEONATES.

15 Inch Colour TFT with touch screen

Alarm function, automatic freezing and ST segment analysis function, 192 HOUR TREND GRAPH, 1000 SETS of NIBP

List, Arrythmia analysis, Single/ Dual IBP,

1. **ECG:** Lead configuration standard 3 leads RA,LA,LL
Lead Mode: I,II,III,
5 leads cable: RA,LA,LL,RL and V
Heart Rate Range 20-230 bpm
Accuracy: $\pm 1\%$ or ± 1 bpm, choosing the greater
Resolution 1 bpm
2. **NIBP:** Measurement Mode : Intelligent Oscillation Mode
Cuff Inflating Time: < 30 sec
Initial Cuff Inflating Pressure: Adult Mode: 180 mmHg
Neonate Mode : 120 mmHg
Maximum Allowed Inflating Pressure : 300 mm Hg
Accuracy of Blood Pressue: Maximum average error ± 5 mm Hg
Protection of Over Pressure : Adult:300 mmHg
Pediatrics: 240mmHg
Neonates: 150mmHg
3. **SPO2:** Sensor: Two-wave LED
Range: 0% -100%
Resolution 1%
Range of Pulse rate : 20 ~ 254 beat/min
Accuracy of pulse rate: ± 1 beat / min
4. **RESP:** Method: Thoracic Impedance
RR Measurement Range: 0 -100 rpm
Accuracy: ± 1 rpm
5. **TEMP:** Measurement Range: 20-45⁰C

AMOUNT in Rupees(IN WORDS).....

Thanking you
Yours truly